



BeckmanCoulter
AU系列培训
DXC700AU 操作介绍
郑振寰

BeckmanCoulter

AU系列培训

DXC700AU 操作介绍

目录


- 1 结构
- 2 每日使用注意事项
- 3 操作界面
- 4 故障

1 结构

DXC700AU的主机结构框架基于AU680。主要变化在于上盖，主上盖分成三个部分。两个试剂盘单独有盖板。主盖板上增加状态指示灯。




1. 状态指示
2. 试剂针臂
3. 光度计单元
4. 搅拌机构
5. 反应盘
6. 样品针臂
7. 进样器
8. 急诊盘
9. 监视器臂
10. 操作按钮
11. 注射器单元


12. ISE单元
13. 清洗剂
14. 稀释清洗齐
15. 水箱
16. 5L清洗剂
17. 试剂仓
18. 上盖



1. 断路器
2. 电源
3. 供水排水管道
4. 黄色诊断开关


红色闪烁-停止
 绿色亮-初始化，热机，测量，重置，结束
 黄色亮-暂停
 蓝色亮-休眠
 关闭-关机






进样器

1. 架条码
2. 条码阅读器窗口
3. 上盖
4. 架装载区
5. 试管架
6. 样品条码阅读器窗口
7. 条码光束



急诊盘

1. 样品杯/试管
2. 急诊盘
3. 小急诊盘盖
4. 条码阅读器窗口
5. 急诊盘旋转LED
6. 急诊开始按钮
7. 诊断/急诊盘旋转按钮
8. 条码阅读器光束



No. 1-No. 22 病人样品、质控品、校准品

S-H ISE高值血清校准品

S-L ISE低值血清校准品

U-H ISE高值尿液校准品

U-L ISE低值尿液校准品

CRS-H ISE高值标准参考溶液（仅日本）

CRS-M ISE中值标准参考溶液（仅日本）


CRS-L ISE低值标准参考溶液（仅日本）

CLEAN ISE清洗剂

SEL-K ISE选择性检查溶液(K)


SEL-Na ISE选择性检查溶液(Na)

RB-1, RB-2 试剂空白




样品针臂

1. 样品针
2. 冲洗池




冲洗针臂

1. 试剂针
2. 冲洗池




搅拌单元

1. R1/S搅拌单元
2. R2搅拌单元
3. 冲洗池




反应盘



反应杯DXC700AU与AU反应杯不同，是5×5mm正方。





光度计灯室



冲洗站

1. 冲洗针
2. 溢流针
3. 吐水针
4. 吸取针
5. 干燥针
6. 干燥棒
7. 抽空针






试剂盘

1. R2试剂仓盖
2. R1试剂仓盖
3. R2试剂仓小盖
4. R1试剂仓小盖
5. R2试剂仓
6. R1试剂仓




注射器

1. R2蓝色注射器
2. R1蓝色注射器
3. ISE缓冲液注射器
4. 黄色样品注射器
5. 蓝色冲洗注射器




前仓试剂

1. 5L清洗剂
2. 水箱
3. 稀释清洗剂
4. 2L清洗剂



清洗剂蠕动泵

1. 蠕动泵管
2. 蠕动泵
3. 接头



ISE单元


1. 搅拌装置
2. 样品杯
3. C1 电极
4. Na 电极
5. K 电极
6. REF 电极
7. 夹断阀
8. 混合样品吸取蠕动泵

9. 中标液吸取蠕动泵




ISE前仓试剂

1. ISE 参比液
2. ISE 中标液
3. ISE 缓冲液
4. ISE 操作LED



控制电脑


1. 键盘
2. 监视器
3. 鼠标
4. 手工条码阅读器
5. 电脑



1. ESC键
2. F1-F8功能键
3. F9开始键
4. F10暂停键
5. F11进样器停止键
6. Print Screen屏幕打印键
7. Scroll Lock菜单列表键
8. Pause事件清除键
9. Num Lock数字小键盘指示
10. 数字小键盘切换
11. F12 停止或休眠键
12. Home 复位键
13. End 关机键
14. 数字键
15. ↑, ↓, ←, →方向键

软件界面

1. 导航菜单
2. 主工作区
3. 蓝色活动按钮
4. 事件显示区域
5. 绿色或白色状态显示
6. 子工作区
7. 返回上一屏幕



主工作区

1. 屏幕按钮
2. 标签页
3. 操作区
4. 功能按钮

<< STAT > STAT (Patient)

STAT Status

STAT (Patient)

STAT (Calibration)

STAT (QC)

Auto ACAL / QC Setup

Test Order

Demographics

Cup Pos.

Sample Kind

STAT

Sample No.

001

Type

Serum

Sample ID

Samples Remaining

9996

Sample Dilution Rate

1

Panel

Change Group Display

Group of Tests

1. Analysis Test

*: Analyzed Test

					Normal	Diluted	Concentrated
ALT	CHOL	GGT19	LDH	GLUC	CRP	ALB	AMY
Cl	K	Na					

Search by Sample ID

Test Total

Select Auto Rerun Orders

Perform Replicates

Selected Tests

0

Sample Volume(uL)

0

(not include dead volume)

Rerun F3

Pending List F4


Batch Entry F5

Delete Test Order F6



Batch Order from LIS F7

Print F8


1. 事件日志按钮
2. 帮助按钮
3. 事件清除按钮



隐藏导航面板



1. 主页
2. 导航面板
3. 相关链接
4. 分层链接
5. 搜索条
6. 打印
7. 高亮显示搜索切换
8. 下一个主题
9. 上一个主题





2 每日使用注意事项

DXC700AU的每日开机流程与AU680完全一致。

2.1 例行检查

检查试剂、清洗剂余量，检查进水状态。

检查可能存在的泄漏。




检查可能
存在的堵塞

①







②




支持的试剂瓶

1. 试剂仓盖
2. 60ml 试剂瓶
3. 15ml 试剂瓶
4. 30ml 试剂瓶
5. 120ml 试剂瓶
6. 适配器
7. 间隔






1. 15 mL瓶
2. 8 mm
3. 30 mL瓶
4. 4 mm
5. 60 mL瓶
6. 12 mm
7. 最高液位
8. 120 mL瓶
9. 10 mm






试管/样品杯


1. 样品针
2. 短样品试管/杯
3. 死腔量
4. 最大样品针下降深度
5. 长样品试管/杯



1. 试管架
2. 条码标签
3. 样品管
4. 光束小于5°
5. 最小7mm间隔



Color	Rack Application	Magnet 1, 2, 3
White 	Used to analyze first-run samples and rerun samples.	● ○ ○
Blue 	Used to calculate reagent blanks for creating calibration curves.	● ○ ●
Yellow 	Used to create calibration curves.	● ● ○
Color	Rack Application	Magnet 1, 2, 3
Green 	Used to analyze control samples.	○ ○ ●
Red 	Used to analyze emergency samples.	○ ● ○



1






2



3




4






3 操作界面

DXC700AU的操作界面与AU680的内容一致，界面布置有所区别。

初始界面





Test Name Parameters

Panel

Group of Tests

Test Name

Common Reagents

LIH Reagent

Dedicated

No.	Test Name	Long Name	Reagent ID	Alarm Tests	Multi Reagent Switch	Reagent Detail
1	ALB	Albumin	002	10	No	
2	ALP	Alkaline Phosphatase	004	10	No	
3	ALT	Alanine Transferase	007	10	No	
4	AMY	Amylase	006	10	No	
5	AST	Aspartate Transferase	009	10	No	
6	CO2	Bicarbonate	037	10	No	
7	DBILC	Direct Bilirubin	011	10	No	Color(Blank:8.DBB)
8	DBB	Direct Bilirubin Blank	010	10	No	Blank(Color:7.DBILC)
9	TBILC	Total Bilirubin	012	10	No	Color(Blank:10.TBB)
10	TBB	Total Bilirubin Blank	025	10	No	Blank(Color:9.TBILC)
11	CA	Calcium Arsenazo	117	10	No	
12	NDBILI	Direct Bili - 181	181	10	No	
13	CHOL	Cholesterol	016	10	No	
14	AMMON	Ammonia	154	10	No	
15				10	No	
16	CK	Creat Phospho Kinase	079	10	No	
17	CRE	Creatinine	078	10	No	
18	GGT	Gamma Glutamyl Trans	019	10	No	
19	GLU	Glucose	021	10	No	
20				10	No	

Edit
F1

Sample Blank
F5

Calculated Tests
F6

Delete
Test Parameters

Print
F8

项目列表

Sample Blank

Color

17DBILC

29TBILC

3None

4None

5None

6None

7None

8None

9None

10None

Blank

8DBB

10TBB

None

None

None

None

None

None

None

None

Close

Calculated Tests

Calculated Tests

189.%SATr

290.LDL-X

391.FR/AL

492.TIBC

5None

6None

7None

8None

9None

10None

11None

12None

13None

14None

15None

16None

17None

18None

19None

20None

Close

<< Configuration Parameters > Common Test Parameters > Test Name Parameters

Test Name Parameters

Panel

Group of Tests

Test Name

Common Reagents

<Editing>

No.	Common Reagent Name	Reagent ID	Alarm Tests	Onboard Stability Period	
				Day	Hour
1			32		
2			32		
3			32		
4			32		
5			32		
6			32		
7			32		
8			32		
9			32		
10			32		

Save

F1

项目面板

病人项目面板

<< Configuration Parameters > Common Test Parameters > Panel

Test Name Parameters		Panel		Group of Tests	
Patient Sample		RB / ACAL		QC	
Type	Serum	Panel Name		0.Default	
Panel Name		Default		Selected Tests 0	
1.ALB	2.ALP	3.ALT	4.AMY	5.AST	6.CO2
7.DBILC	8.DBB	9.TBILC	10.TBB	11.CA	12.NDBILI
13.CHOL	14.AMMON	15.	16.CK	17.CRE	18.GGT
19.GLU	20.	21.IRON	22.	23.LDH	24.
25.LIPASE	26.Li	27.MG	28.PHOS	29.TP	30.eCHOL
31.eBUN	32.BUN	33.TRIG	34.	35.	36.
37.UA	38.UCSFP	39.	40.	41.HDL-G	42.UBCr
43.FRUC	44.LDL-G	45.BHB	46.	47.	48.
49.	50.FERR	51.	52.	53.	54.
55.	56.	57.	58.	59.MALB	60.PALB
61.RF	62.TRF	63.	64.	65.	66.ACETA
67.	68.AMIK	69.CARB	70.DIG	71.GENT	72.
73.	74.	75.	76.	77.PBARB	78.PHENY
79.	80.	81.	82.SAL	83.THEO	84.TOP
85.VPA	86.VANC	87.	88.	89.%SATr	90.LDL-X
91.FR/AL	92.TIBC	93.TIBCr	94.	95.	96.LIH
97.Na	98.K	99.Cl	100.HbA1c%	101.T-Hb	102.HbA1c
103.	104.	105.	106.	107.	108.
109.	110.	111.	112.	113.	114.
115.	116.	117.	118.	119.	120.

Edit F1
Print F8

试剂空白 校准品面板

Test Name Parameters

Panel

Group of Tests

Patient Sample

RB / ACAL

QC

Panel Name

0.

Panel Name

Type

Serum

Selected Tests

51

1.ALB	2.ALP	3.ALT	4.AMY	5.AST	6.CO2	7.DBILC	8.DBB
9.TBILC	10.TBB	11.CA	12.NDBILI	13.CHOL	14.AMMON	15.	16.CK
17.CRE	18.GGT	19.GLU	20.	21.IRON	22.	23.LDH	24.
25.LIPASE	26.Li	27.MG	28.PHOS	29.TP	30.eCHOL	31.eBUN	32.BUN
33.TRIG	34.	35.	36.	37.UA	38.UCSFP	39.	40.
41.HDL-G	42.UIBC _r	43.FRUC	44.LDL-G	45.BHB	46.	47.	48.
49.	50.FERR	51.	52.	53.	54.	55.	56.
57.	58.	59.MALB	60.PALB	61.RF	62.TRF	63.	64.
65.	66.ACETA	67.	68.AMIK	69.CARB	70.DIG	71.GENT	72.
73.	74.	75.	76.	77.PBARB	78.PHENY	79.	80.
81.	82.SAL	83.THEO	84.TOP	85.VPA	86.VANC	87.	88.
89.%SAT _r	90.LDL-X	91.FR/AL	92.TIBC	93.TIBCr	94.	95.	96.LIH
97.Na	98.K	99.Cl	100.HbA1c%	101.T-Hb	102.HbA1c	103.	104.
105.	106.	107.	108.	109.	110.	111.	112.
113.	114.	115.	116.	117.	118.	119.	120.

ACAL+RB

One Point

RB Only

Edit

F1

Print

F8

Test Name Parameters

Panel

Group of Tests

Patient Sample

RB / ACAL

QC

Panel Name

0.

Panel Name

Type

Serum

Selected Tests

0

1.ALB	2.ALP	3.ALT	4.AMY	5.AST	6.CO2	7.DBILC	8.DBB
9.TBILC	10.TBB	11.CA	12.NDBILI	13.CHOL	14.AMMON	15.	16.CK
17.CRE	18.GGT	19.GLU	20.	21.IRON	22.	23.LDH	24.
25.LIPASE	26.Li	27.MG	28.PHOS	29.TP	30.eCHOL	31.eBUN	32.BUN
33.TRIG	34.	35.	36.	37.UA	38.UCSFP	39.	40.
41.HDL-G	42.UIBCr	43.FRUC	44.LDL-G	45.BHB	46.	47.	48.
49.	50.FERR	51.	52.	53.	54.	55.	56.
57.	58.	59.MALB	60.PALB	61.RF	62.TRF	63.	64.
65.	66.ACETA	67.	68.AMIK	69.CARB	70.DIG	71.GENT	72.
73.	74.	75.	76.	77.PBARB	78.PHENY	79.	80.
81.	82.SAL	83.THEO	84.TOP	85.VPA	86.VANC	87.	88.
89.%SATr	90.LDL-X	91.FR/AL	92.TIBC	93.TIBCr	94.	95.	96.LIH
97.Na	98.K	99.Cl	100.HbA1c%	101.T-Hb	102.HbA1c	103.	104.
105.	106.	107.	108.	109.	110.	111.	112.
113.	114.	115.	116.	117.	118.	119.	120.

Edit
F1

Print
F8



Analysis Mode								System Setup								Program the Logon								Comment Master								User Menu							
Inspect Analyzer Status																Review QC Data Charts																							
Analyzer Daily Maintenance Perform Routine Inspections																Monitor Results Status, Time to Completion, Flags																							
Monitor Reagent Status Perform Reagent Volume Check																																							
Inspect ISE Reagents Review Number of Tests and Stability																																							
ISE Daily Maintenance Clean and Calibrate the ISE																																							
Order and Perform Calibration from Racks																																							
Review Calibration Data Recent and History																																							
Order and Perform QC from Racks																																							
Edit F1																																							

Calibration Monitor		ISE Maintenance		Calibration Setup		
Calibrators		General		ISE		
☐ Bar Code Operation						
No.	Cup Position	Type	Calibrator		Lot No.	Expiration
			Name	ID		
1	0001-01	Serum	Cal Level 1	DR0070-1	6102K11	03/31/2016
2	0001-02	Serum	Cal Level 2	DR0070-2	6102K11	03/31/2016
3	0001-03	Serum	CO2 20	2340-C	59722875	01/31/2016
4	0001-04	Serum	CO2 40	2340-E	59722877	01/31/2016
5	0001-05	Serum	HDL CAL	GENZYME 80-6449-00	803091	07/05/2016
6	0001-06	Serum	FRUCTOSAMINE CAL	FSC	159243-02	07/13/2016
7	0001-07	Serum	LIPASE CAL	IN KIT	2321	05/01/2016
8	0001-08	Serum	LDL CALIBRATOR	GENZYME 80-5666-02	805885	08/09/2016
9	0001-09	Serum	BETA-HYDROXYBUTYRATE	BHB	113071CE	08/31/2016
10	0001-10	Serum				
11	0002-01	Serum				
12	0002-02	Serum				

<RB Sample Information>

No.	Name	ID	Serum	Urine	Other-1	Other-2
1	DI Water		☒	☒	☒	☒
2			☐	☐	☐	☐

Edit F1	Calibrator Info. Loading History F3	Set Conc Value F5	Print F8
------------	---	----------------------	-------------

原厂校准品条码扫描登记界面

Load Calibrator Information

Step 1:

Scan the 2D bar code after selecting the following edit box.
If there is more than one bar code, scan all bar codes continuously.
Don't touch the keyboard during scanning.

The calibrator information scanned is displayed below.

Catalog No. Expiration
Lot No. Ver.

Level	Calibrator Name	Calibrator ID
1		
2		
3		
4		
5		
6		
7		

Step 2:

Select the position that calibrator is set, and then press "Set Parameters" button.
In the case of multi-level calibrator, it is continuously set from the selected position.

1.




Type

☐ ID Update

Set Parameters

Close

Load Calibrator Information

Step 1:

Scan the 2D bar code after selecting the following edit box.
If there is more than one bar code, scan all bar codes continuously.
Don't touch the keyboard during scanning.

The calibrator information scanned is displayed below.

Catalog No. Expiration
Lot No. Ver.

Level	Calibrator Name	Calibrator ID
1	SYSTEM CALIBRATOR	6630010118
2		
3		
4		
5		
6		
7		

Step 2:

Select the position that calibrator is set, and then press "Set Parameters" button.
In the case of multi-level calibrator, it is continuously set from the selected position.

1.




Type

☐ ID Update

Set Parameters

Close

校准品浓度设置

校准品历史信息

Calibrator Info. Loading History

Calibrator

1 AU Multi Calibrator

Performed

User Name

Calibrator Name

Catalog No.

Level

Lot No.

Expiration

Ver.

Close

36

Set Conc Value

Calibrator

1 AU Multi Calibrator

Type

Serum

Test Name	Point	Conc.	Test Name	Point	Conc.	Test Name	Point	Conc.
1.AST	1	10	2.ALT	1	1	3.LDH		
4.rGTP	1	20	5.AMY			6.TG		
7.TCHO			8.HDL			9.LDL		
10.			11.			12.		
13.			14.			15.		
16.			17.			18.		
19.			20.			21.		
22.			23.			24.		
25.			26.			27.		
28.			29.			30.		
31.			32.			33.		
34.			35.			36.		
37.			38.			39.		
40.			41.			42.		
43.			44.			45.		
46.			47.			48.		
49.			50 A/G			51 ALB	1	10
52 GLB	1	10	53.			54.		
55.			56.			57.		
58.			59.			60.		

Close

Calibration Monitor

ISE Maintenance

Calibration Setup

Status

RB History

RB Detail

Calibration History

Calibration Detail

Type

Serum

 Display

All Tests

<Status> : All tests passed : Some tests Expire Soon : Some tests have failed

	Serum	Urine	Other-1	Other-2
Reagent Blank				
Calibration				

: Impossible to judge because of reagent unchecked.

Test Name		Reagent Blank	Calibration	Test Name		Reagent Blank	Calibration
1.ALB	*			2.ALP	*		
3.ALT	*			4.AMY	*		
5.AST	*			32.BUN	*		
11.CA	*			13.CHOL	*		
16.CK	*			99.CI			
6.CO2	*			17.CRE	*		
7.DBILC	*			43.FRUC	*		
18.GGT	*			19.GLU	*		
41.HDL-G	*			21.IRON	*		
98.K				23.LDH	*		

* : Advanced Calibration

Factor List
F4

RB/CAL
Selection
F6

1

2

RB/CAL Selection

Type Serum

Test Name	Seq.	R1(R1-1)		R2(R2-1)		R1-2		Reagent Blank	Calibration
		Lot No.	Bottle No.	Lot No.	Bottle No.	Lot No.	Bottle No.		
11.TP	1	7801	0570	7801	0581			Expired	Expired
18.CO2								No Data	No Data
15.AST	1	7589	2790	7589	2806			Passed RB	
8.BUN	1	8113	3593	8113	3611			Passed RB	Passed Calibration
9.GGT	1	7628	1491	7628	1508			Passed RB	
10.LDH	1	7582	1820	7582	1780			Passed RB	
12.ALP	1	7645	0141	7645	0206			Passed RB	
13.AMY	1	7545	1968					Passed RB	
13.AMY	2	8275	2330					Passed RB	
14.ALT	1	7617	0908	7617	0919			Passed RB	



Calibration Monitor		ISE Calibration		Calibration Setup	
Status		RB History		RB Detail	
Calibration History		Calibration Detail			

Test Name 6.CO2 ↶ ↷

Date/Time 1.06/17/2016 09:21 Passed

Type Serum

Reagent Lot No. Bottle No.

R1(R1-1) 0001 0001

R2(R2-1)

R1-2

Sequence 1

RB Expiration Date 06/18/2016 09:21

Method END1

Wavelength 340 nm nm

Point-1 0 27

Point-2


Reagent Blank P0 1.0257 Px Py 1.0257

	OD		OD		OD		OD
P0	1.0257	P7	1.0257	P14	1.0257	P21	1.0257
P1	1.0257	P8	1.0257	P15	1.0257	P22	1.0257
P2	1.0257	P9	1.0257	P16	1.0257	P23	1.0257
P3	1.0257	P10	1.0257	P17	1.0257	P24	1.0257
P4	1.0257	P11	1.0257	P18	1.0257	P25	1.0257
P5	1.0257	P12	1.0257	P19	1.0257	P26	1.0257
P6	1.0257	P13	1.0257	P20	1.0257	P27	1.0257

--	--	--	--

Comment Select

	RB/CAL Selection F2	Data Select F3			Flags Contents F6		Print F8
--	---------------------------	-------------------	--	--	-------------------------	--	-------------



Calibration Monitor

ISE Calibration

Calibration Setup

Status

RB History

RB Detail

Calibration History

Calibration Detail

Test Name

6.CO2

Type

Serum

Date/Time

1.06/16/2016 22:03

Passed

Reagent

Lot No.

Bottle No.

R1(R1-1)

0001

0001

R2(R2-1)

R1-2

Sequence

1

Cal Expiration Date

06/17/2016 22:03

Reagent Blank

Cal Type

2AB

Formula

Polygonal

Factor

A0=1.4749E002

B0=0.0000E000

A1=1.6090E002

B1=1.1300E-002

OD

Previous Curve

0.4000

0.3200

0.2400

0.1600

0.0800

0.0000

0

8

16

24

32

40

	Cal No.	Conc.	OD	Flags
1	3	20	0.1356	
2	4	40	0.2599	

Conc.

Comment

Select

RB/CAL Selection

F2

Data Select

F3

Recalculation

F4

Flags Contents

F6

Graph Scale

F7

Print

F8



样品命令

病人样品架

Rack
(Patient)

Rack
(Calibration)

Rack
(QC)

Test Order

Demographics

Sample Kind

Routine

Switch

Sample No.

0004

Type

Serum

Sample ID

Samples Remaining

9996

Sample Dilution Rate

1

Panel

☐ Change Group Display

Group of Tests

1.Analysis Test

* : Analyzed Test

ALT	CHOL	GGT19	LDH	GLUC	Normal	Diluted	Concentrated
CI	K	Na					

Search by Sample ID

Test Total

Select
Auto Rerun Orders

Perform Replicates

Selected Tests

0

Sample Volume(uL)

0

(not include dead volume)

Rerun
F3

Pending List
F4

Batch Entry
F5

Delete Test Order
F6

Batch Order
from LIS
F7

Print
F8

<< STAT > STAT Status

44

STAT Status			STAT (Patient)		STAT (Calibration)		STAT (QC)		Auto ACAL / QC Setup	
					Calibrator				Control	
Pos.	Kind	Cup	Status	Sample No.	Type	Cup Information		ID		Comment
1	Patient	Present		001	Serum			412316		
2	Patient	Present		002	Serum			123412		
3	Patient	None								
4	Patient	None								
5	Patient	None								
6	Patient	None								
7	Patient	None								
8	Patient	None								
9	Patient	None								
10	Patient	None								
11	Patient	None								
12	Patient	None								
13	Patient	None								
14	Patient	None								
15	Patient	None								
16	Patient	None								
17	Patient	None								
18	Patient	None								
19	Patient	None								
20	Patient	None								
21	Patient	None								
22	Patient	None								

RB Cup Information

Pos.	Cup	Status	Cup Information	Comment
RB1	Present			
RB2	Present			Cup Exists

Delete Sample Assignment

STAT Pause
F2

STAT Check
F3

Display Sample
Assignment
F4

CAL/QC Position
F5

STAT Table
Setting
F6

Previous Cup Set
F7

Sample ID Edit

Sample Manager		Reaction Monitor		Data Statistics		Correlation Chart	
Main		By Patient Sample		By Patient Test			
Index	2.10/19/2015 11:06		-	2.10/19/2015 11:06			
Select All Samples		Clear All Samples		☐ Abnormal ☐ Rerun Data			
No.[^]	Sample No.	Cup Position	Sample ID	Run Date/Time	Status		
39	0029	0072-04	ODC0014	10/19/2015 13:54:03	Done		
40	0030	0072-05	ODC0014	10/19/2015 13:54:07	Done		
41	0031	0072-06	ODC0014	10/19/2015 13:54:12	Done		
42	0032	0072-07	ODC0014	10/19/2015 13:54:16	Done		
43	0033	0072-08	ODC0014	10/19/2015 13:54:21	Done		
44	0034	0072-09	ODC0014	10/19/2015 13:54:26	Done		
45	0035	0072-10	ODC0014	10/19/2015 13:54:30	Done		
46	0036	0056-01	ODC0015	10/19/2015 13:54:35	Done		
47	0037	0056-02	ODC0015	10/19/2015 13:54:39	Done		
48	0038	0056-03	ODC0015	10/19/2015 13:54:44	Done		
49	0039	0056-04	ODC0015	10/19/2015 13:54:48	Done		
50	0040	0056-05	ODC0015	10/19/2015 13:54:53	Done		
51	0041	0056-06	ODC0015	10/19/2015 13:54:57	Done		
52	0042	0056-07	ODC0015	10/19/2015 13:55:02	Done		
53	0043	0056-08	ODC0015	10/19/2015 13:55:06	Done		
54	0044	0056-09	ODC0015	10/19/2015 13:55:11	Done		
55	0045	0056-10	ODC0015	10/19/2015 13:55:15	Done		

	Copy to Disk F2	Find F3	External Memory Remove F4	Remove External Memory F4	Data Correction F6	Send to LIS F7	Print F8
--	--------------------	------------	---------------------------------	---------------------------------	--------------------------	-------------------	-------------

Sample Status

Detail

Realtime Display

Sample Manager

ISE Calibration

View mode

Static

Sample No.	Cup Position	Sample ID	Order	Status	Results	
0308	0006-08		17:02	Done		
0309	0006-09		17:02	Done		
0310	0006-10		17:03	Done		
0311	0009-01		17:03	Done		
0312	0009-02		17:03	Done		
0313	0009-03		17:04	Done		
0314	0009-04		17:04	Done		
0315	0009-05		17:05	Done		
0316	0009-06		17:05	Done		
0317	0009-07		17:05	Done		
0318	0009-08		17:06	Done		
0319	0009-09		17:06	Done		
0320	0009-10		17:06	Done		

Find
F2

Rerun
F3

Switch to
Static View

Reaction
Monitor
F6



[illegible]

Sample Status		Detail		Realtime Display		Sample Manager		ISE Calibration	
All		Quick							
						View mode		Static	
Sample No.0311	Serum	Sample ID							
d1.0-1	5867	d1.0-2	6128	d1.0-3	6024	d1.0-4	6000		
d1.0-5	6048								
Sample No.0312	Serum	Sample ID							
d1.0-1	6045	d1.0-2	5988	d1.0-3	6029	d1.0-4	6108		
d1.0-5	6008								
Sample No.0313	Serum	Sample ID							
d1.0-1	6047	d1.0-2	6050	d1.0-3	6014	d1.0-4	5981		
d1.0-5	6070								
Sample No.0314	Serum	Sample ID							
d1.0-1	6065	d1.0-2	6026	d1.0-3	6087	d1.0-4	6035		
d1.0-5	6052								
Sample No.0315	Serum	Sample ID							
d1.0-1	6010	d1.0-2	6006	d1.0-3	6029	d1.0-4	6060		
d1.0-5	6052								
Sample No.0316	Serum	Sample ID							
d1.0-1	6115	d1.0-2	6105	d1.0-3	6116	d1.0-4	6069		
d1.0-5	5969								
Sample No.0317	Serum	Sample ID							
d1.0-1	5941	d1.0-2	6003	d1.0-3	5983	d1.0-4	5945		
d1.0-5	6104								
Sample No.0318	Serum	Sample ID							
d1.0-1	6066	d1.0-2	5995	d1.0-3	5991	d1.0-4	5889		
d1.0-5	6033								
Sample No.0319	Serum	Sample ID							
d1.0-1	6080	d1.0-2	5963	d1.0-3	6013	d1.0-4	6049		
d1.0-5	5996								
Sample No.0320	Serum	Sample ID							
d1.0-1	6158	d1.0-2	6013	d1.0-3	6027	d1.0-4	5926		
d1.0-5	6116								
		Find	Switch to Static View		Reaction Monitor				
		F2			F6				

Sample Manager

Reaction Monitor

Data Statistics

Correlation Chart

Main

General

ISE

Index2.08/27/2015 23:06Sample No.0001TypeSerum

SampleRoutineQC/Cal. No. Seq. No. Counts100003 / 00066

Test Name1.ALBID6024011829Cuvette No.63

	R1(R1-1)	Sample	R2(R2-1)	Preprocess
Mix Bar No.	2	3	2	
CONC	Result			
Reaction OD	4			
	0.6184			

LIH Measuring Point

Flag

	Reagent Information	Lot No.	Bottle No.	Pos.
Sample	R1(R1-1)	2884	1457	39
	R2(R2-1)			
Cal.	R1(R1-1)	2884	1457	39
	R2(R2-1)			
RB	R1(R1-1)	2884	1457	39
	R2(R2-1)			

	Date/Time
Measured	08/28/2015 00:24:13
Calibration	08/27/2015 19:17:22
Reagent Blank	08/27/2015 19:14:42
Photocal	08/27/2015 18:14:51

	600nm	800nm	Reaction OD	Reagent Blank
0	0.2500	0.0024	0.2476	0.2478
1	0.8819	0.0035	0.8784	0.2602
2	0.8950	0.0034	0.8916	0.2624
3	0.9045	0.0034	0.9011	0.2634
4	0.9118	0.0036	0.9082	0.2637
5	0.9178	0.0036	0.9142	0.2638
6	0.9217	0.0035	0.9182	0.2637
7	0.9244	0.0034	0.9210	0.2637
8	0.9268	0.0034	0.9234	0.2637
9	0.9295	0.0036	0.9259	0.2638
10	0.9318	0.0034	0.9284	0.2638
11	0.9259	0.0035	0.9224	0.2612
12	0.9280	0.0033	0.9247	0.2622
13	0.9296	0.0033	0.9263	0.2626

Max OD	0.9445	0.0036	0.9409	0.2638
Min OD	0.2500	0.0024	0.2476	0.2478
Reaction OD	0.6184			
Photocal	0.4902	0.5583		

Copy to DiskF2

List DisplayF3

Remove External MemoryF4

Chart ViewF5

PrintF8

List Display

Range of Date 10/19/2015 13:39 - 10/19/2015 13:55

Number of Samples 45

No.	Cup Position	Sample No.	Sample ID	Run Date/Time	Test Name				
1	0027-01	0001	water	10/19/2015 13:40:24	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC	15.CRP		
2	0027-02	0002	water	10/19/2015 13:41:00	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC	15.CRP		
3	0027-03	0003	water	10/19/2015 13:41:36	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC	15.CRP		
4	0027-04	0004	water	10/19/2015 13:42:12	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC	15.CRP		
5	0027-05	0005	water	10/19/2015 13:42:49	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC	15.CRP		
6	0002-01	0006	ODC0003	10/19/2015 13:43:43	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC			
7	0002-02	0007	ODC0003	10/19/2015 13:44:15	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC			
8	0002-03	0008	ODC0003	10/19/2015 13:44:46	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC			
9	0002-04	0009	ODC0003	10/19/2015 13:45:18	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC			
10	0002-05	0010	ODC0003	10/19/2015 13:45:50	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC			
11	0002-06	0011	ODC0003	10/19/2015 13:46:21	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC			
12	0002-07	0012	ODC0003	10/19/2015 13:46:53	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH
					13.UN	14.GLUC			
13	0002-08	0013	ODC0003	10/19/2015 13:47:25	8.ALB	9.ALT	10.CHOL	11.GGT19	12.LDH

Print

Output

Close

Sample Manager

Reaction Monitor

Data Statistics

Correlation Chart

Main

General

Index

1.09/22/2015 16:23

-

1.09/22/2015 16:23

Test Name

All

Cuvette No.

*

Preprocess Cuvette No.

*

☐ Abnormal

Patient

RB/Cal./QC

Mix Bar No. R11

*

Sample

*

R21

*

Preprocess

*

(Preprocess:Pre-dilution/Preprocess)

Sample Kind	Search by Sample No.						QC/Cal. No.	Search by Control / Calibrator ID
Calibration		-		*	-	*	* *	
QC		-		*	-	*	* *	
RB	R001	-	R001	*	-	*	*	

Counts

Sequence

Calibration

☒ 1
☒ 2
☒ 3
☒ 4

☒ 1
☒ 2
☒ 3
☒ 4
☒ 5

QC

☒ 1
☒ 2
☒ 3
☒ 4


☒ 1
☒ 2
☒ 3
☒ 4
☒ 5

RB

☒ 1
☒ 2
☒ 3
☒ 4

☒ 1
☒ 2
☒ 3
☒ 4
☒ 5

Search Condition F5



反应监测 数据

Sample Manager

Reaction Monitor

Data Statistics

Correlation Chart

Main

General

ISE

Index Sample No. Type

Sample QC/Cal. No. Seq No. Counts

Test Name ID

R1(R1-1)

Sample

R2(R2-1)

Preprocess

Mix Bar No.

2

3

2

Result

4

0.6184

LIH Measuring Point

Flag

Reagent Information

Lot No.

Bottle No.

Pos.

Sample

R1(R1-1)

2884

1457

39

Cal.

R1(R1-1)

2884

1457

39

RB

R1(R1-1)

2884

1457

39

Date/Time

Measured

08/28/2015 00:24:13

Calibration

08/27/2015 19:17:22

Reagent Blank

08/27/2015 19:14:42

Photocal

08/27/2015 18:14:51

600nm

800nm

Reaction OD

Reagent Blank

0

0.2500

0.0024

0.2476

0.2478

1

0.8819

0.0035

0.8784

0.2602

2

0.8950

0.0034

0.8916

0.2624

3

0.9045

0.0034

0.9011

0.2634

4

0.9118

0.0036

0.9082

0.2637

5

0.9178

0.0036

0.9142

0.2638

6

0.9217

0.0035

0.9182

0.2637

7

0.9244

0.0034

0.9210

0.2637

8

0.9268

0.0034

0.9234

0.2637

9

0.9295

0.0036

0.9259

0.2638

10

0.9318

0.0034

0.9284

0.2638

11

0.9259

0.0035

0.9224

0.2612

12

0.9280

0.0033

0.9247

0.2622

13

0.9296

0.0033

0.9263

0.2626

Max OD

0.9445

0.0036

0.9409

0.2638

Min OD

0.2500

0.0024

0.2476

0.2478

Reaction OD

0.6184

Photocal

0.4902

0.5583

Copy to Disk

List Display

Remove External Memory

Chart View

Print

F2

F3

F4

F5

F8

Sample Manager		Reaction Monitor		Data Statistics		Correlation Chart			
Main		Statistics		Chart View		Data View		Histogram	
Index		12.08/24/2015 17:52		-		7.08/26/2015 22:26			
☒ Search all samples ☐ Search the designated sample									
Search by Sample ID									
Sample Kind		Search by Sample No.				Search by Sample ID			
☒ Routine		-		*	-	*	*		
☒ Emergency		-		*	-	*	*		
☐ STAT		-		*	-	*	*		
Type		☒ Serum		☒ Urine		☒ Other-1		☒ Other-2	
						Search by Patient Info. F5			

Sample Manager

Reaction Monitor

Data Statistics

Correlation Chart

Main

Statistics

Chart View

Data View

Histogram

Index

12.08/24/2015 17:52

-

7.08/26/2015 22:26

Sample No.

*

-

*

Type

(All)

Sample Kind

(All)

Sample ID

*

Number of Samples

4023

Other Type

All

Sex

All

Age(Years/Months)

0

/

0

-

150

/

11


Test Name[^]	Number of Data	Mean	SD	CV(%)	Range	Max.	Min.
1.ALB	169	4.07	0.559	13.73	3.2	5.3	2.1
2.ALP	289	92.96	75.736	81.47	715.4	749.0	33.6
3.ALT	1826	26.2	25.99	99.20	390	393	3
4.AMY	17	44.8	20.14	44.96	77	100	23
5.AST	578	29.9	37.40	125.08	467	474	7
6.CO2	326	26.7	2.90	10.86	21	36	15
7.DBILC	31	0.1	0.34	340.00	1	1	0
9.TBILC	238	0.7	0.87	124.29	6	6	0
11.CA	439	9.5	0.59	6.21	3	11	8
13.CHOL	1814	185.0	41.15	22.24	409	495	86
16.CK	142	176.4	304.19	172.44	2863	2884	21
17.CRE	2145	1.1	0.60	54.55	12	12	0


Print
F8



Sample Manager		Reaction Monitor		Data Statistics		Correlation Chart	
Main		Statistics		Chart View		Data View	
Index		12.08/24/2015 17:52 - 7.08/26/2015 22:26		Sample No. * - *		Type (All)	
Sample Kind (All)		Sample ID *		Number of Samples 4023		Other Type All	
Sex All		Age(Years/Months) 0 / 0 - 150 / 11					

No.	S.No.	Measured Time	1 ALB	3 ALT	4 AMY
1	0001	08/24/2015 19:27		12 r	
2	0003	08/24/2015 19:27		11 r	
3	0004	08/24/2015 19:27		16 r	
4	0005	08/24/2015 19:28		16 r	
5	0006	08/24/2015 19:28		11 r	
6	0007	08/24/2015 19:28		14 r	
7	0009	08/24/2015 19:28		14 r	
8	0010	08/24/2015 19:29		32 r	
9	0011	08/24/2015 19:27		15 r	
10	0014	08/24/2015 19:27		16 r	
11	0016	08/24/2015 19:28		14 r	
12	0018	08/24/2015 19:28		18 r	
13	0019	08/24/2015 19:29		15 r	
14	0020	08/24/2015 19:29		18 r	
15	0022	08/24/2015 19:29		13 r	
16	0023	08/24/2015 19:29		4 r	
17	0024	08/24/2015 19:29		19 r	
18	0026	08/24/2015 19:30	4.3 r	33 r	





Sample Manager

Reaction Monitor

Data Statistics

Correlation Chart

Main

Data View

Chart View

Index - Sample No. - Type
Sample Kind Sample ID Number of Samples Other Type
Sex Age(Years/Months) / - /

No.	S.No.	Measured Time	X Axis 1.ALB				Y Axis 2.ALP			
1	0026	08/24/2015 19:30	4.3	r			72.4	r		
2	0161	08/24/2015 19:41	3.6	r			79.5	r		
3	0162	08/24/2015 19:41	3.9	r			54.9	r		
4	0247	08/24/2015 19:51	4.2	r			51.4	r		
5	0280	08/24/2015 19:55	3.9	r			46.7	r		
6	0310	08/24/2015 20:26	4.1	r			83.9	r		
7	0562	08/24/2015 21:00	4.2	r			80.2	r		
8	0659	08/24/2015 21:09	4.1	r			108.8	r		
9	0649	08/24/2015 21:09	3.0	r			53.5	r		
10	0661	08/24/2015 21:10	3.6	r			68.3	r		
11	0650	08/24/2015 21:10	4.0	r			36.1	r		
12	0666	08/24/2015 21:11	4.2	r			50.3	r		
13	0669	08/24/2015 21:12	3.8	r			106.7	r		
14	0693	08/24/2015 21:15	4.2	r			79.3	r		
15	0696	08/24/2015 21:16	4.3	r			113.2	r		
16	0709	08/24/2015 21:17	3.4	r			109.2	r		
17	0718	08/24/2015 21:20	4.6	r			78.6	r		
18	0729	08/24/2015 21:20	4.3	r			151.5	r		

☐ : Deselected from Chart View

Select / Deselect
F5

试剂界面

试剂管理

ISE

<< Reagent > Reagent Management

Reagent Management
Reagent Inventory
Reagent Consumption
STAT
(Calibration)
Rack
(Calibration)

Main
Details

ISE

Accessible

READY

ISE Maintenance

Bottle Name	Cell	Serum Tests	Urine Tests	Onboard Remaining	Expiration	Lot No.	Comment
REF Solution	1	> 408	> 300	71D		00001	
MID Solution	1	> 123	> 48	71D		00001	
BUF Solution	1	> 94	> 66	71D		00001	

Serum

Urine

Cal Stability

Cal Stability

Comment

Comment

Electrode Lot No.

Ref. Electrode Lot No.

Cell 1

Na

K

Cl

Replace ISE Reagent
F1

ISE Parameter Setting
F4

Reagent Check
F5

Print
F8

yeee

Reagent Management
Reagent Inventory
Reagent Consumption
STAT
(Calibration)
Rack
(Calibration)

Auto

Manual

Index

-

Type

Serum

Decision

Auto

Tests Dispensed

R1 Volume

R2 Volume

* : 3- Part Reagent Analysis

Test Name	Reagent ID	Monday	Tuesday	Wednesday	Thursday	Friday	Saturday	Sunday
1.W1		----	----	----	----	----	----	----
2.W2		----	----	----	----	----	----	----
3.W3		----	----	----	----	----	----	----
4.R-150		----	----	----	----	----	----	----
5.S-1.0	111	----	----	----	----	----	----	----
6.S-1.1		----	----	----	----	----	----	----
7.S-1.6		----	----	----	----	----	----	----
8.BUN	034	0	0	0	0	0	0	0
9.GGT	019	0	0	0	0	0	0	0
10.LDH	027	0	0	0	0	0	0	0
11.TP	032	0	0	0	0	0	0	0
12.ALP	004	0	0	0	0	0	0	0
13.AMY	006	0	0	0	0	0	0	0
14.ALT	007	0	0	0	0	0	0	0
15.AST	009	0	0	0	0	0	0	0
16.CK	079	0	0	0	0	0	0	0
17.CREA	078	0	0	0	0	0	0	0
18.CO2	037	0	0	0	0	0	0	0
19.S1.0P	111	----	----	----	----	----	----	----
20.GLU	021	0	0	0	0	0	0	0

Display Range
F3

Change
Reagent Type
F5

Print
F8

Reagent Management		Reagent Inventory		Reagent Consumption		STAT (Calibration)		Rack (Calibration)		
Date	1	2	-	3	4	Type	5			
Tests Dispensed		R1 Volume		R2 Volume		Reportable Tests		Cumulative Tests Dispensed		
* : 3- Part Reagent Analysis										
Test Name	Reagent ID	Routine	Emergency	STAT	Sub Total	Rerun	RB	ACAL	QC	Total
1.ALB	002	5763	11	0	5774	134	58	46	1151	7163
2.ALP	004	8416	32	0	8448	23	375	0	1043	9889
3.ALT	007	24393	7	0	24400	15	482	0	1209	26106
4.AMY	006	2226	61	0	2287	55	542	0	1294	4178
5.AST	009	9297	7	0	9304	5	420	0	1099	10828
6.CO2	037	9748	7	0	9755	29	410	795	1108	12097
7.DBILC	011	1470	26	0	1496	266	62	62	1021	2907
8.DBB	010	1470	26	0	1496	266	62	62	1021	2907
9.TBILC	012	6183	7	0	6190	14	28	28	1111	7371
10.TBB	025	6181	7	0	6188	14	28	28	1111	7369
11.CA	117	6765	7	0	6772	16	48	92	1062	7990
12.NDBILI	181	0	0	0	0	0	0	0	0	0
13.CHOL	016	20747	4	0	20751	43	34	30	1101	21959
14.AMMON	154	0	0	0	0	0	0	0	0	0
15.		0	0	0	0	0	0	0	0	0
16.CK	079	3171	8	0	3179	12	432	0	1177	4800
17.CRE	078	29631	24	0	29655	176	518	1020	1279	32648
18.GGT	019	691	6	0	697	15	166	0	287	1165
19.GLU	021	22950	5	0	22955	100	40	36	1170	24301
20.		0	0	0	0	0	0	0	0	0
Total		353352	632	0	353984	2114	5942	3567	35005	400612
Samples		-----	-----	-----	-----	-----	-----	-----	-----	-----
Copy to Disk F2		Display Range F3		Remove External Memory F4		Change Reagent Type F5				Print F8

事件日志

Date Range		12/24/2015 10:47:17 - 12/25/2015 10:47:17		Count	35
No.	Date/Time	Level	Event No.	Event	
1	12/25/2015 10:47:01	3	5005	MEASURE START	
2	12/25/2015 10:46:19	3	5020	MEASURE END	
3	12/25/2015 10:45:13	3	5005	MEASURE START	
4	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (13)	
5	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (12)	
6	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (11)	
7	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (10)	
8	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (9)	
9	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (8)	
10	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (7)	
11	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (6)	
12	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (5)	
13	12/25/2015 10:44:41	2	6001	STAT BAR CODE READ ERROR (4)	

Copy to Disk
F2


Filter
F3

Remove
External Memory
F4


Details
F6

Refresh
F7

Print
F8




1




2

日间 质控图

<< Quality Control > Chart





Chart

Twin Plot Chart

QC Setup

Main

Chart View

Data Review

Index

9.08/25/2015 22:09

Type

Serum

QC Sample No.

Q001

Test Name	Level	Control Name	Lot No.	Result	Flags				SD Range
1.ALB /1	1	1.Biorad Liquicheck 1	16641	2.7	r				*
1.ALB /1	2	2.Biorad Liquicheck 2	16642	4.2	r				*
2.ALP /1	1	1.Biorad Liquicheck 1	16641	93.1	r				*
2.ALP /1	2	2.Biorad Liquicheck 2	16642	406.1	r				*
3.ALT /1	1	1.Biorad Liquicheck 1	16641	26	r				*
3.ALT /1	2	2.Biorad Liquicheck 2	16642	100	r				1
4.AMY /1	1	1.Biorad Liquicheck 1	16641	60	r				*
4.AMY /1	2	2.Biorad Liquicheck 2	16642	328	r				1
5.AST /1	1	1.Biorad Liquicheck 1	16641	28	r				*
5.AST /1	2	2.Biorad Liquicheck 2	16642	162	r				*
32.BUN /1	1	1.Biorad Liquicheck 1	16641	15	r				*
32.BUN /1	2	2.Biorad Liquicheck 2	16642	50	r				*
11.CA /1	1	1.Biorad Liquicheck 1	16641	8	r				*
11.CA /1	2	2.Biorad Liquicheck 2	16642	12	r				*
13.CHOL /1	1	1.Biorad Liquicheck 1	16641	113	r				1
13.CHOL /1	2	2.Biorad Liquicheck 2	16642	257	r				1
16.CK /1	1	1.Biorad Liquicheck 1	16641	114	r				2
16.CK /1	2	2.Biorad Liquicheck 2	16642	416	r				1
99.CI	1	1.Biorad Liquicheck 1	16641	87	r				*
99.CI	2	2.Biorad Liquicheck 2	16642	101	r				*

: Omitted from the statistics calculation target

Edit

Index Comment

Test Comment

Test

Detail

F1

F2

F3

F5

F6

分析仪 维护界面

<< Maintenance > Analyzer Maintenance

Analyzer Maintenance						ISE Maintenance		File Management		Version Information	
Maintenance			Consumables		Photocal Monitor						
☐ Display Pending Items Only ☐ Procedures Due Soon											
No. [^]	Maintenance	Help	Pending (Number of Times)	Performed	Frequency	Update		Approval			
1	Clean the Sample Probe (1)	?	12/25/2015 14:14	12/18/2015 14:14	Weekly	☒ Analyzer Maintenance					
2	Clean the Mix Bars (0)	?	12/25/2015 14:16	12/18/2015 14:16	Weekly	Replace Sample Probe	Replace Sample Syringe				
3	Perform a W2 (0)	?	12/25/2015 14:30	12/18/2015 14:30	Weekly	Clean Sample Probe	Replace Reagent Probe/Syringe				
4	Perform a Photocal (1)	?	12/25/2015 15:00	12/18/2015 15:00	Weekly	Replace Wash Syringe	Clean Wash Well				
5	Clean the Pre-dilution Bottle (0)	?	12/25/2015 14:17	12/18/2015 14:17	Weekly	Replace Wash Nozzle	Prime Washing Line				
6	Clean the Sample Probe Wash Well (0)	?	01/18/2016 14:17	12/18/2015 14:17	Monthly	Prime Wash Nozzle	Replace Mix Bar				
7	Clean the Reagent Probe Wash Wells (0)	?	01/18/2016 14:17	12/18/2015 14:17	Monthly			Replace Roller Pump tubing			
8	Clean the Mix Bar Wash Wells (0)	?	01/18/2016 14:17	12/18/2015 14:17	Monthly						
9	Clean the Wash Nozzle and Inspect the Tube Mounting Joints (0)	?	01/18/2016 14:17	12/18/2015 14:17	Monthly						
10	Clean the Deionized Water Filter (0)	?	01/18/2016 14:17	12/18/2015 14:17	Monthly						
(0): No Action (1): Run Control (2): Calibrate and Run Control											
Maintenance Edit		Maintenance History		Maintenance Schedule		Stand By		W1 F5		W2 F6	
								Photocal F7		Print F8	

1

2

Analyzer Maintenance

ISE Maintenance

File Management

Version Information

Maintenance

Calibration

Selectivity Check

Measurement

Display Pending Items Only

Procedures Due Soon

No. [^]	Maintenance	Help	Pending (Number of Times)	Performed	Frequency
1	Clean the ISE (2)		01/15/2016 14:17	01/14/2016 14:17	Daily
2	Serum Calibration (1)		01/15/2016 14:17	01/14/2016 14:17	Daily
3	Urine Calibration (1)		01/15/2016 14:17	01/14/2016 14:17	Daily
4	Selectivity Check for the Na/K Electrodes (1)		12/25/2015 14:16	12/18/2015 14:16	Weekly
5	Enhanced Cleaning of Electrode Line (2)		12/25/2015 14:30	12/18/2015 14:30	Weekly
6	Manually Clean the ISE Sample Pot and Sample Pot Tubing (2)		01/01/2016 15:00	12/18/2015 15:00	2Weeks
7	Manually Clean the ISE Mix Bar and Liquid Level Sensors (0)		01/01/2016 15:00	12/18/2015 15:00	2Weeks
8	Inspect and Add ISE Internal Reference Solution (2)		02/18/2016 14:17	12/18/2015 14:17	2Months
9	Replace the Roller Pump Tubing (2)		03/18/2016 14:17	12/18/2015 14:17	3Months
10	Replace the Tubing around the Electrode Block (2)		03/18/2016 14:17	12/18/2015 14:17	3Months

Update

Approval

ISE Maintenance

Total Prime

Prime Bypass

Buffer Prime

MID/REF Prime

Drain Flowcell

Drain Bypass

Cleaning

Cleaning (Enhanced)

(0): No Action

(1): Run Control

(2): Calibrate and Run Control

Maintenance Edit

Maintenance History

Maintenance Schedule

Print

Analyzer Maintenance			ISE Maintenance			File Management			Version Information		
Maintenance			Calibration			Selectivity Check			Measurement		
ISE Status			READY			Electrode Status			Na K Cl		
						Reagent Status			REF MID Buffer		

< Result >

Date/Time	Serum						Urine					
	Slope			MID Solution Factor			Slope			MID Solution Factor		
	Na	K	Cl	Na	K	Cl	Na	K	Cl	Na	K	Cl
Last Result	54.6	54.9	-54.2	1.000	1.000	0.900	54.5	55.0	-55.5	1.100	1.000	1.100
12/18/2015 11:11	54.6	54.9	-54.2	1.000	1.000	0.900	54.5	55.0	-55.5	1.100	1.000	1.100
12/18/2015 11:09							54.0	54.2	-55.5	1.200	1.100	1.100
12/18/2015 11:07	53.0	55.0	-54.0	1.100	1.000	1.100						
12/18/2015 10:46	55.0	55.0	-55.0	1.100	0.900	1.200	54.0	54.0	-56.0	1.200	1.200	1.100

< Acceptable Calibration Range >

High	68.0	68.0	-38.0	1.200	1.200	1.200	68.0	68.0	-38.0	1.200	1.200	1.200
Low	38.0	38.0	-68.0	0.800	0.800	0.800	38.0	38.0	-68.0	0.800	0.800	0.800

Serum Start			Urine Start			Serum/Urine Start		
-------------	--	--	-------------	--	--	-------------------	--	--

	ISE Stop F4	Graph Display F5	Detail F6		Print F8
--	----------------	---------------------	--------------	--	-------------

Maintenance Schedule

Year/Month 2016/05


☐ ISE☒ Performed☐ Pending☐ Overdue

Frequency	Target	Maintenance	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
Weekly	ANL	Clean the Sample Probe																															
Weekly	ANL	Clean the Mix Bars																															
Weekly	ANL	Perform a W2																															
Weekly	ANL	Perform a Photocal																															
Weekly	ANL	Clean the Pre-dilution Bottle																															
Monthly	ANL	Clean the Sample Probe Wash Well																															
Monthly	ANL	Clean the Reagent Probe Wash Wells																															
Monthly	ANL	Clean the Mix Bar Wash Wells																															
Monthly	ANL	Clean the Wash Nozzle and Inspect the Tube Mounting Joints																															
Monthly	ANL	Clean the Deionized Water Filter																															

OK

Analyzer Maintenance		ISE Maintenance		File Management		Version Information																													
Maintenance		Calibration		Selectivity Check		CRS Calibration		Measurement																											
ISE Status		READY		Electrode Status		Na		K		Cl		Reagent Status		REF		MID		Buffer																	
Type		Serum		Switch		Date Range		12/17/2015 16:12		-		12/27/2015 23:12																							
Slope Na K Cl <table border="1"><thead><tr><th colspan="2">Min Slope</th><th colspan="2">Max Slope</th></tr></thead><tbody><tr><td>Na</td><td>40.3</td><td>Na</td><td>47.0</td></tr><tr><td>K</td><td>39.2</td><td>K</td><td>44.8</td></tr><tr><td>Cl</td><td>-50.6</td><td>Cl</td><td>-54.4</td></tr></tbody></table>																				Min Slope		Max Slope		Na	40.3	Na	47.0	K	39.2	K	44.8	Cl	-50.6	Cl	-54.4
Min Slope		Max Slope																																	
Na	40.3	Na	47.0																																
K	39.2	K	44.8																																
Cl	-50.6	Cl	-54.4																																
Serum Start				Urine Start				Serum/Urine Start																											
								ISE Stop F4				Data Display F5				Detail F6				ISE Monitor				Print F8											

[illegible]



Analyzer Maintenance

ISE Maintenance

File Management

Version Information

Operation

☐ Save Files to HD

☐ Load Files from HD

☒ Save Files to External Media

☐ Load Files from External Media

<File List>

File
Test Name Parameters
Panel
Group of Tests
Test Volume and Methods
Rerun Test Parameters
Rerun Check Parameters
Checked Tests
Contamination Parameters
Data Check Parameters
Analysis Mode
System Setup
Program the Logon
Comment Master
User Menu

Copy to / from Disk
F2

Remove
External Memory
F4

File Select
F6

Copy to Disk

File save is successful.

Test Name Parameters		
Panel		
Group of Tests		
Test Volume and Methods		
Rerun Test Parameters		
Rerun Check Parameters		
Checked Tests		
Contamination Parameters		
Data Check Parameters		
Online		
Sample Program Format		
List Format		
Analysis Mode		
System Setup		
Program the Logon		
Comment Master		
User Menu		
Calibration Parameters		
QC Parameters		
STAT Status		
Auto ACAL / QC Setup		

OK

Analyzer Maintenance	ISE Maintenance	File Management	Version Information
----------------------	-----------------	-----------------	---------------------

System / Console

Program	Ver.	Rev.
System	1	04
Console	1	25
Help	1	00
-----	--	--

Analyzer


Program	Ver.	Rev.	Station	Program	Ver.	Rev.	Station
Analyzer IPL ROM			---	Rack Back, Transfer Controller(CF,CG)	3	2	31
Analyzer Control Program	01	09	---	Temperature Controller (Incubation,Fan)	3	1	28
Primary Station Firmware	3	4	0	Temperature Controller (Wash and DI Water)	3	1	27
Reagent Refrigerator Controller(DA01)	3	2	2	Dilution Pump Controller	3	2	9
Reagent Refrigerator Controller(DA11)	3	2	3	ISE Data Controller	3	5	13
STAT Table Controller(DC)	3	2	8	ISE Mixer Controller	3	2	15
Sample Probe Controller(FA)	3	2	7	ISE Pump Controller	3	2	14
Cuvette Washer Controller(FB)	3	2	29	-----	--	--	---
Mixer Controller(FC01)	3	2	26	-----	--	--	---
Mixer Controller(FC11)	3	2	6	-----	--	--	---
R1 Probe Controller(FD01)	3	2	5	-----	--	--	---
R2 Probe Controller(FD11)	3	2	4	-----	--	--	---
Photometry Controller(GA)	2	1	200	-----	--	--	---
Sample Dispenser Controller(SA)	3	2	10	-----	--	--	---
R1 Dispenser Controller(SA)	3	2	11	-----	--	--	---
R2 Dispenser Controller(SA)	3	2	12	-----	--	--	---
Rack Transfer Controller(CB,CJ)	3	2	32	-----	--	--	---
Rack Extrusive Controller(CC)	3	2	33	-----	--	--	---
Normal Lane Controller(CD)	3	2	34	-----	--	--	---
Buffer, Repeat Controller(CE,CH)	3	2	30	-----	--	--	---

Print
F8

4 故障

DXC700AU的数据旗标与AU系列完全一样。

标记	定义	解释
d	用户剔除质控结果	用户手动将 QC 数据从计算中剔除。
e	用户编辑数据	数据已被编辑。
(	清洗液短缺, 导致污染参数异常	机内、在仓或在板清洗剂缺少, 防交叉污染清洗停止, 有这个旗标可能会发生携带污染。
Wa	已用错误的比色杯分析结果	已使用未达到光电校准标准的比色杯分析结果
R	检测到试剂不足	液面探测显示试剂量不足
#	检测到样品不足	样品体积不足或液面探测出现故障。
%	检测到堵塞	在吸取样品时样品针堵塞或样品凝结。
?	无法计算结果	速率反应中读数少于三个, 比色杯漫盘, 发生机械错误。
M	当 AU5800 与 Beckman 实验室自动化系统 连接时, 样品 ID 的读取与当前 在轨道线或 AU5800 上处理中的样品 ID 相同。	
n	没有进行 LIH 测试	检查样品, 必要时重测, 检查 LIH 试剂。
l	结果可能受高脂血症影响	结果可能受到高血脂的影响或者样品极其浑浊
i	结果可能受黄疸影响	结果可能受黄疸影响。
h	结果可能受溶血影响	结果可能受溶血影响。
Y	试剂空白OD高于最后点OD设置	试剂空白OD高于为最后点光度计读取点设定的试剂OD限制范围。
U	试剂空白OD低于最后点OD设置	试剂空白OD低于为最后点光度计读取点设定的试剂OD限制范围。
y	试剂空白/常规OD第高于初始点OD限制	试剂空白的初始读点OD或正常分析的P0处OD高于初始读点设定的试剂OD限制范围。
u	试剂空白/常规OD第低于初始点OD限制	试剂空白的初始读点OD或正常分析的P0处OD低于初始读点设定的试剂OD限制范围。
@	OD 高于 3.0	异常高的值。反应OD超出3.0。在双波长测定中, 如果其中任意一个波长超出3.0 OD, 则发生错误。



AU系列培训

DXC700AU 操作介绍

\$	没有足够的数据来确定反应线性	速率反应中在指定的可接受光密度范围内的读数点少于三个。为正确计算速率反应, 在达到最高或最低光密度前必须获得三个读数。如果超出这些光密度限制, 反应可能因为高浓度或试剂状况的问题导致底物耗竭。当发生这种错误时, 不进行线性计算。
D	反应的 OD 值高于最高 OD 范围	指定读数点的光密度超出速率和固定检测法中的吸光度范围。 - 指定读数点 FST+2, 即第一个光度测定点加 2 的光密度 (上升反应速率法) - 指定读数点 LST+2, 即最后一个光度测定点加 2 的光密度 (下降反应速率法) - 第一个光度测定点的光密度 (固定法)
B	反应的 OD 值低于最低 OD 范围	在上升反应或下降 (更可能发生) 反应速率法中, 光度测定点 FST+2 (第一个光度测定点加 2) 的光密度没有达到最低光密度值。
*	速率法线性错误	当反应速率的变动幅度超出测试参数定义的%变动范围时, 生成这个标记, 因此被视为非线性。
&	前带测试数据异常	用于前带判定的数据异常。
Z	前带错误	满足逻辑检验 1、2 或 3 中任意一个数据检查方程式。这通常是由于样品中分析物的浓度异常高造成的。
E	在速率法分析中检测到过度反应	在速率法检测中, 如果反应在很短的时间内完成, 则被视为过度反应, 结果被判定为错误。这通常是由于样品中分析物的浓度异常高造成的。
Fx	结果 (OD) 高于动态范围	无法计算浓度。样品的 OD 超出动态范围的上限 OD。
Gx	结果 (OD) 低于动态范围	无法计算浓度。样品的 OD 低于动态范围的下限 OD。
!	无法计算浓度	系统无法计算结果。检查反应监测数据。
)	样品分析时使用的试剂批号与用于 RB/ 校准的批号不同	试剂批号与校准的试剂批号不匹配。
a	试剂已过期	试剂过期或者在线时间超出具体测试参数中限定的时间。

4 故障

ba	没有校准数据或过期	没有试剂空白或校准数据，或者数据过期。
bh	没有使用最新的校准/RB	最近的试剂空白或校准失效或过期。结果可能错误，不应报告。
bn	使用了主曲线(拟合曲线)	没有进行校准或者校准不成功。系统已使用主曲线生成结果。
bz	使用前带数据的校准曲线	系统已经使用前带数据的校准曲线生成结果。这与真实的结果相似，应该用来估算重测的稀释率。
F	结果高于动态范围	样品浓度高于测试参数设定的动态范围上限。
G	结果低于动态范围	样品浓度低于测试参数中设定的动态范围的下限，或者没有正确进行吸取试剂。
ph	结果高于上危急值	结果高于上危急值上限。
pl	结果低于下危急值	结果低于上危急值下限。
T	在内部化学检查中发现异常	已检查的测试结果超出菜单列表>参数>其他>已检查的测试中指定的检查范围。
P	阳性	定性结果：样品结果超出上限。
N	阴性	定性结果：样品结果低于下限。
H	结果高于参考范围	样品结果高于正常范围中设定的具体范围的上限。
L	结果低于参考范围	样品结果低于正常范围中设定的具体范围的下限。
J	结果高于重测判定范围	结果高于重测判定范围。
K	结果低于重测判定范围	结果低于重测判定范围。
fh	结果高于重测反射范围	结果高于用户指定的反射范围。
fl	结果低于重测反射范围	结果低于用户指定的反射范围。
Va	多次测量检查的偏差超出范围	试剂空白或校准的重复精确性超出允许的范围
xQ	多规则质控在一个质控中检测到失败	如果使用多次检查质控规则的两对数据中有一对超出范围，数据的另一部分被标记。
1Q	QC数据超出在“单次检查水平”字段中输入的范围	QC 数据的某一点超出特定质控的检查选项卡上的单次检查水平中定义的 SD。
2Q	QC 数据超出 13s 质控范围	QC 数据的某一点超出质控具体参数的检查选项卡上的“多次检查水平”中定义的 ±3SD。
3Q	QC 数据超出 22s 质控范围	两个邻近的 QC 数据点超出同一方向的 ±2SD 质控限值。

4Q	QC 数据超出 R4s 质控范围	两个连续的高和低浓度 QC 数据点中的一个超出 +2SD 的质控限值，另一个超出 -2SD 的质控限值，或两个质控之间的差距超出 4SD。
5Q	QC 数据超出 41s 质控范围	四个连续的 QC 数据点结果超出 1SD 质控限制。
6Q	预设数目的连续的质控结果偏向均值一侧	预设数目（7-10）的连续数据点结果全部大于或小于均值。
7Q	连续的质控结果反映了持续上升或下降的值	预设数目（4-10）的连续数据点结果持续增加或下降。
S	样品重测，重测结果替代原始结果	重复运行了样品测试，并用此重测结果替换了先前结果成为最终结果。
/	待测或未分析	尽管被要求，没有进行测试（通常是因为缺少试剂），或测试正在进行中。
r	结果已经通过在线通信传送到 LIS（主机）	
c	用户校准了结果	尽管被要求，没有进行测试（通常是因为缺少试剂），或测试正在进行中。



yeec维修网



yeec维修网淘宝店铺



yeec维修网微信公众号



yeec维修网QQ群



yeec维修网站长QQ



yeec维修网站长微信



BeckmanCoulter

AU系列培训

DXC700AU 操作介绍

本讲义归属权为原作者，由<http://www.yeec.com>压缩整理，未经许可不得传播或用于商业用途